

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
 Data File : VN052787.D
 Acq On : 11 Dec 2018 7:37
 Operator : MD\SY
 Sample : J6316-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 51 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 AVE-X-B-BALL-FIELD-S-3

Quant Time: Dec 11 10:49:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1260785	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1954841	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1728082	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	695567	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	647250	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
35) Dibromofluoromethane	7.59	113	548425	45.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.02%	
50) Toluene-d8	10.09	98	2399729	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
62) 4-Bromofluorobenzene	12.40	95	767818	48.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.42%	
Target Compounds						
16) Acetone	3.82	43	30632	8.460	ug/l	94
20) Methylene Chloride	4.56	84	35578	2.511	ug/l	95
43) Isopropyl Acetate	8.17	43	77172	4.048	ug/l #	72
59) 2-Hexanone	10.76	43	273975	38.161	ug/l #	52

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121018\
Data File : VN052787.D
Acq On : 11 Dec 2018 7:37
Operator : MD\SY
Sample : J6316-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 51 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
AVE-X-B-BALL-FIELD-S-3

Quant Time: Dec 11 10:49:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

