

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
 Data File : VN052762.D
 Acq On : 10 Dec 2018 19:10
 Operator : MD\SY
 Sample : PB115448ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115448ZHE#11

Quant Time: Dec 11 02:59:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1195725	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1854252	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1672764	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	683126	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	624840	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
35) Dibromofluoromethane	7.59	113	496304	43.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.84%	
50) Toluene-d8	10.09	98	2272667	51.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.14%	
62) 4-Bromofluorobenzene	12.40	95	741846	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
Target Compounds						
16) Acetone	3.82	43	11235	3.272	ug/l	100
20) Methylene Chloride	4.56	84	22481	1.673	ug/l	94
43) Isopropyl Acetate	8.17	43	80458	4.449	ug/l #	72


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121018\
Data File : VN052762.D
Acq On : 10 Dec 2018 19:10
Operator : MD\SY
Sample : PB115448ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115448ZHE#11

Quant Time: Dec 11 02:59:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN052762.D

Acq: 10 Dec 2018 19:10

Instrument :

MSVOA_N

ClientSampleId :


PB115448ZHE#11

Tot Ion:168 Resp: 1195725

Ion Ratio Lower Upper

168 100

99 50.9 41.7 62.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN0

7.67

600000

500000

400000



300000

200000

100000

0

Time-->



#16

Acetone

Concen: 3.272 ug/l

RT: 3.82 min Scan# 636

Delta R.T. 0.00 min

Lab File: VN052762.D


Acq: 10 Dec 2018 19:10

Tgt Ion: 43 Resp: 11235

Ion Ratio Lower Upper

43 100

58 32.3 26.0 39.0



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0

3.82

5000

4000


3000


2000

1000

0

Time-->





#20

Methylene Chloride

Concen: 1.673 ug/l

RT: 4.56 min Scan# 867

Delta R.T. 0.01 min

Lab File: VN052762.D

Acq: 10 Dec 2018 19:10

Instrument :

MSVOA_N

ClientSampleId :

PB115448ZHE#11

Tot Ion: 84 Resp: 22481


Ion Ratio Lower Upper

84 100

49 118.2 98.0 147.0

51 30.7 30.1 45.1

86 57.4 50.7 76.1






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



#33

1,2-Dichloroethane-d4

Concen: Below ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN052762.D


Acq: 10 Dec 2018 19:10

Tgt Ion: 65 Resp: 624840

Ion Ratio Lower Upper



65 100


67 55.5 0.0 109.8



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.59 min Scan# 2119
Delta R.T. 0.00 min
Lab File: VN052762.D
Acq: 10 Dec 2018 19:10

Instrument :
MSVOA_N
ClientSampleId :
PB115448ZHE#11

Tot Ion:114 Resp: 1854252
Ion Ratio Lower Upper
114 100
63 20.2 0.0 41.8
88 15.7 0.0 31.6




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): V
Ion 87.90 (87.60 to 88.60): V






#35
Dibromofluoromethane
Concen: 50.000 ug/l
RT: 7.59 min Scan# 1810
Delta R.T. 0.00 min
Lab File: VN052762.D
Acq: 10 Dec 2018 19:10

Tgt Ion:113 Resp: 496304
Ion Ratio Lower Upper
113 100
111 103.5 82.0 123.0
192 22.2 17.0 25.6



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 4.449 ug/l

RT: 8.17 min Scan# 1989

Delta R.T. 0.01 min

Lab File: VN052762.D

Acq: 10 Dec 2018 19:10

Instrument :

MSVOA_N

ClientSampleId :

PB115448ZHE#11


Tot Ion: 43 Resp: 80458

Ion Ratio Lower Upper

43 100

61 12.6 24.1 36.1#




87 6.5 10.6 15.8#



Abundance Ion 43.00 (42.70 to 43.70): VN052762.D (-1968) (-)

Ion 61.00 (60.70 to 61.70): VN052762.D (-1968) (-)

Ion 87.00 (86.70 to 87.70): VN052762.D (-1968) (-)



#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN052762.D


Acq: 10 Dec 2018 19:10

Tgt Ion: 98 Resp: 2272667

Ion Ratio Lower Upper



98 100

100 64.5 51.3 76.9



Abundance Ion 98.00 (97.70 to 98.70): VN052762.D (-2570) (-)

Ion 100.00 (99.70 to 100.70): VN052762.D (-2570) (-)





#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 12.40 min Scan# 3305

Delta R.T. 0.00 min

Lab File: VN052762.D

Acq: 10 Dec 2018 19:10

Instrument :

MSVOA_N

ClientSampleId :

PB115448ZHE#11





Tot Ion: 95 Resp: 741846

Ion Ratio Lower Upper

95 100

174 88.8 0.0 173.0

176 86.3 0.0 166.2



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2997

Delta R.T. 0.00 min

Lab File: VN052762.D

Acq: 10 Dec 2018 19:10




Tgt Ion: 117 Resp: 1672764


Ion Ratio Lower Upper

117 100

82 53.5 43.6 65.4

119 31.8 25.4 38.2





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN052762.D

Acq: 10 Dec 2018 19:10

Instrument :

MSVOA_N

ClientSampleId :

PB115448ZHE#11


Tot Ion:152 Resp: 683126

Ion Ratio Lower Upper

152 100

115 56.4 28.4 85.2

150 157.8 0.0 346.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

