

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121018\
 Data File : VN052784.D
 Acq On : 11 Dec 2018 6:18
 Operator : MD\SY
 Sample : J6316-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 48 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 BORING-SAMPLE-CBH-591

Quant Time: Dec 11 07:33:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1192698	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1849690	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1614096	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	634279	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	626008	48.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.42%	
35) Dibromofluoromethane	7.59	113	535790	46.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.98%	
50) Toluene-d8	10.09	98	2234732	50.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.70%	
62) 4-Bromofluorobenzene	12.40	95	704595	47.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.48%	
Target Compounds						
16) Acetone	3.82	43	18345	5.356	ug/l	97
20) Methylene Chloride	4.56	84	33216	2.478	ug/l	96
43) Isopropyl Acetate	8.17	43	79087	4.384	ug/l #	74


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121018\
Data File : VN052784.D
Acq On : 11 Dec 2018 6:18
Operator : MD\SY
Sample : J6316-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 48 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BORING-SAMPLE-CBH-591

Quant Time: Dec 11 07:33:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. -0.00 min

Lab File: VN052784.D

Acq: 11 Dec 2018 6:18

Instrument :

MSVOA_N

ClientSampleId :


BORING-SAMPLE-CBH-591

Tot Ion:168 Resp: 1192698

Ion Ratio Lower Upper




168 100

99 50.9 41.7 62.5



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#16

Acetone

Concen: 5.356 ug/l

RT: 3.82 min Scan# 638

Delta R.T. 0.01 min

Lab File: VN052784.D


Acq: 11 Dec 2018 6:18

Tgt Ion: 43 Resp: 18345

Ion Ratio Lower Upper



43 100


58 34.3 26.0 39.0



Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 58.00 (57.70 to 58.70): VNO





#20

Methylene Chloride

Concen: 2.478 ug/l

RT: 4.56 min Scan# 866

Delta R.T. 0.00 min

Lab File: VN052784.D

Acq: 11 Dec 2018 6:18

Instrument :

MSVOA_N

ClientSampleId :

BORING-SAMPLE-CBH-591

Tot Ion: 84 Resp: 33216


Ion Ratio Lower Upper

84 100

49 116.0 98.0 147.0

51 37.9 30.1 45.1

86 61.4 50.7 76.1






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



#33

1,2-Dichloroethane-d4

Concen: N.D. ug/l

RT: 8.03 min Scan# 1945

Delta R.T. -0.00 min

Lab File: VN052784.D


Acq: 11 Dec 2018 6:18

Tgt Ion: 65 Resp: 626008

Ion Ratio Lower Upper



65 100


67 54.4 0.0 109.8



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.59 min Scan# 2119

Delta R.T. -0.00 min

Lab File: VN052784.D

Acq: 11 Dec 2018 6:18

Instrument :

MSVOA_N

ClientSampleId :

BORING-SAMPLE-CBH-591


Tot Ion:114 Resp: 1849690

Ion Ratio Lower Upper

114 100

63 20.1 0.0 41.8




88 15.2 0.0 31.6



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN

Ion 87.90 (87.60 to 88.60): VN



#35

Dibromofluoromethane

Concen: 50.000 ug/l

RT: 7.59 min Scan# 1810

Delta R.T. -0.00 min

Lab File: VN052784.D

Acq: 11 Dec 2018 6:18


Tgt Ion:113 Resp: 535790

Ion Ratio Lower Upper

113 100

111 102.4 82.0 123.0



192 22.1 17.0 25.6




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 4.384 ug/l

RT: 8.17 min Scan# 1989

Delta R.T. 0.01 min

Lab File: VN052784.D

Acq: 11 Dec 2018 6:18

Instrument :

MSVOA_N

ClientSampleId :

BORING-SAMPLE-CBH-591


Tot Ion: 43 Resp: 79087

Ion Ratio Lower Upper

43 100

61 13.0 24.1 36.1#




87 8.0 10.6 15.8#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587

Delta R.T. -0.00 min

Lab File: VN052784.D


Acq: 11 Dec 2018 6:18

Tgt Ion: 98 Resp: 2234732

Ion Ratio Lower Upper



98 100


100 64.6 51.3 76.9



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0





#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN052784.D

Acq: 11 Dec 2018 6:18

Instrument :

MSVOA_N

ClientSampleId :

BORING-SAMPLE-CBH-591


Tot Ion: 95 Resp: 704595

Ion Ratio Lower Upper

95 100

174 89.3 0.0 173.0




176 86.7 0.0 166.2



Abundance Ion 94.90 (94.60 to 95.60): VN052784.D (-3291) (-)

Ion 173.80 (173.50 to 174.50): VN052784.D (-3291) (-)

Ion 175.80 (175.50 to 176.50): VN052784.D (-3291) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2997

Delta R.T. -0.00 min

Lab File: VN052784.D

Acq: 11 Dec 2018 6:18


Tgt Ion: 117 Resp: 1614096

Ion Ratio Lower Upper

117 100

82 53.2 43.6 65.4



119 32.5 25.4 38.2




Abundance Ion 116.90 (116.60 to 117.60): VN052784.D (-2980) (-)

Ion 82.00 (81.70 to 82.70): VN052784.D (-2980) (-)

Ion 118.90 (118.60 to 119.60): VN052784.D (-2980) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. -0.00 min

Lab File: VN052784.D

Acq: 11 Dec 2018 6:18

Instrument :

MSVOA_N

ClientSampleId :

BORING-SAMPLE-CBH-591


Tot Ion:152 Resp: 634279

Ion Ratio Lower Upper

152 100

115 55.5 28.4 85.2

150 157.3 0.0 346.0



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

