

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121019\
 Data File : VN059539.D
 Acq On : 10 Dec 2019 12:02
 Operator : JC/SP
 Sample : VN1210WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1210WBL01

Quant Time: Dec 11 06:35:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 10 01:09:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	251384	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	403513	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	436177	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	142904	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.01	65	177623	50.56	ug/l	0.00
Spiked Amount						
			Recovery	=	101.12%	
35) Dibromofluoromethane	7.57	113	127670	53.02	ug/l	0.00
Spiked Amount						
			Recovery	=	106.04%	
50) Toluene-d8	10.08	98	541326	55.07	ug/l	0.00
Spiked Amount						
			Recovery	=	110.14%	
62) 4-Bromofluorobenzene	12.40	95	202444	55.63	ug/l	0.00
Spiked Amount						
			Recovery	=	111.26%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121019\
Data File : VN059539.D
Acq On : 10 Dec 2019 12:02
Operator : JC/SP
Sample : VN1210WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1210WBL01

Quant Time: Dec 11 06:35:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120919W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 10 01:09:52 2019
Response via : Initial Calibration

