

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121019\
 Data File : VN059543.D
 Acq On : 10 Dec 2019 13:42
 Operator : JC/SP
 Sample : K6177-10ME
 Misc : 3.60g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EP-8ME

Quant Time: Dec 11 07:25:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 10 01:09:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	239281	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	382886	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	362008	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	166670	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	168966	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
35) Dibromofluoromethane	7.57	113	117024	51.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.44%	
50) Toluene-d8	10.08	98	467413	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
62) 4-Bromofluorobenzene	12.40	95	163985	47.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.98%	
Target Compounds						
68) m/p-Xylenes	11.62	106	2063	0.403	ug/l	92
69) o-Xylene	11.95	106	1266	0.266	ug/l	85
80) 1,3,5-Trimethylbenzene	12.73	105	2392	0.223	ug/l	90
84) 1,2,4-Trimethylbenzene	13.04	105	6179	0.584	ug/l	100
95) Naphthalene	15.13	128	4169	0.383	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121019\
Data File : VN059543.D
Acq On : 10 Dec 2019 13:42
Operator : JC/SP
Sample : K6177-10ME
Misc : 3.60g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EP-8ME

Quant Time: Dec 11 07:25:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120919W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 10 01:09:52 2019
Response via : Initial Calibration

