

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121019\
 Data File : VN059545.D
 Acq On : 10 Dec 2019 14:26
 Operator : JC/SP
 Sample : K6204-14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 K072-AA-WC-1

Quant Time: Dec 11 07:12:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 10 01:09:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	260355	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	419769	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	385177	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	160262	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	205336	56.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.86%	
35) Dibromofluoromethane	7.57	113	133239	53.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.38%	
50) Toluene-d8	10.08	98	528303	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
62) 4-Bromofluorobenzene	12.40	95	177937	47.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.00%	
Target Compounds						
16) Acetone	3.79	43	15026	8.840	ug/l	92
20) Methylene Chloride	4.52	84	9788	2.938	ug/l	97
43) Isopropyl Acetate	8.15	43	70951	9.544	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121019\
Data File : VN059545.D
Acq On : 10 Dec 2019 14:26
Operator : JC/SP
Sample : K6204-14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
K072-AA-WC-1

Quant Time: Dec 11 07:12:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120919W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 10 01:09:52 2019
Response via : Initial Calibration

