

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121020\
 Data File : VN065053.D
 Acq On : 10 Dec 2020 19:43
 Operator : JC/MD
 Sample : PB133501ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133501ZHE#01

Quant Time: Dec 11 04:11:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 10 16:04:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	159213	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	289732	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	316573	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	128430	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	102855	54.26	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	108.52%
35) Dibromofluoromethane	7.55	113	94158	49.12	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.24%
50) Toluene-d8	10.06	98	381693	51.91	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.82%
62) 4-Bromofluorobenzene	12.38	95	140113	55.59	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.18%

Target Compounds

					Qvalue
16) Acetone	3.79	43	5892	8.763	ug/l 95
20) Methylene Chloride	4.51	84	6826	3.985	ug/l 96
43) Isopropyl Acetate	8.13	43	14656	3.924	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121020\
Data File : VN065053.D
Acq On : 10 Dec 2020 19:43
Operator : JC/MD
Sample : PB133501ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133501ZHE#01

Quant Time: Dec 11 04:11:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 10 16:04:22 2020
Response via : Initial Calibration

