

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121020\
 Data File : VN065054.D
 Acq On : 10 Dec 2020 20:07
 Operator : JC/MD
 Sample : PB133501ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133501ZHE#02

Quant Time: Dec 11 04:13:18 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 10 16:04:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	159159	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	291143	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	319846	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	127247	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	104502	55.15	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.30%
35) Dibromofluoromethane	7.55	113	95428	49.54	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.08%
50) Toluene-d8	10.06	98	384913	52.09	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.18%
62) 4-Bromofluorobenzene	12.38	95	140941	55.65	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	111.30%

Target Compounds

					Qvalue
16) Acetone	3.78	43	5934	8.829 ug/l	92
20) Methylene Chloride	4.50	84	6596	3.852 ug/l	93
43) Isopropyl Acetate	8.13	43	14940	3.981 ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121020\
Data File : VN065054.D
Acq On : 10 Dec 2020 20:07
Operator : JC/MD
Sample : PB133501ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133501ZHE#02

Quant Time: Dec 11 04:13:18 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 10 16:04:22 2020
Response via : Initial Calibration

