

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121020\
 Data File : VN065056.D
 Acq On : 10 Dec 2020 20:56
 Operator : JC/MD
 Sample : PB133501ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133501ZHE#04

Quant Time: Dec 11 04:25:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 10 16:04:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	160658	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	300186	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	332581	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	132239	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	107430	56.16	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	112.32%		
35) Dibromofluoromethane	7.55	113	96992	48.84	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.68%		
50) Toluene-d8	10.06	98	388736	51.02	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	102.04%		
62) 4-Bromofluorobenzene	12.38	95	143398	54.91	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	109.82%		

Target Compounds

					Qvalue
16) Acetone	3.78	43	5730	8.446 ug/l	95
20) Methylene Chloride	4.50	84	6617	3.829 ug/l	96
43) Isopropyl Acetate	8.13	43	15527	4.012 ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121020\
Data File : VN065056.D
Acq On : 10 Dec 2020 20:56
Operator : JC/MD
Sample : PB133501ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133501ZHE#04

Quant Time: Dec 11 04:25:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 10 16:04:22 2020
Response via : Initial Calibration

