

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121020\
 Data File : VN065062.D
 Acq On : 11 Dec 2020 2:12
 Operator : JC/MD
 Sample : PB133501ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133501ZHE#10

Quant Time: Dec 11 04:40:26 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 10 16:04:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	151548	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	296323	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	316994	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	127780	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	7.99	65	104004	57.64	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	115.28%	
35) Dibromofluoromethane	7.55	113	94158	48.03	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	96.06%	
50) Toluene-d8	10.06	98	376300	50.03	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	100.06%	
62) 4-Bromofluorobenzene	12.38	95	139007	53.92	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	107.84%	

Target Compounds

					Qvalue
16) Acetone	3.78	43	17075	26.681 ug/l	94
20) Methylene Chloride	4.51	84	54311	33.313 ug/l	98
43) Isopropyl Acetate	8.13	43	14981	3.922 ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121020\
Data File : VN065062.D
Acq On : 11 Dec 2020 2:12
Operator : JC/MD
Sample : PB133501ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133501ZHE#10

Quant Time: Dec 11 04:40:26 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 10 16:04:22 2020
Response via : Initial Calibration

