

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121020\
 Data File : VN065064.D
 Acq On : 11 Dec 2020 3:00
 Operator : JC/MD
 Sample : PB133501ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133501ZHE#12

Quant Time: Dec 11 04:44:43 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 10 16:04:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	164133	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	311868	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	334182	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136365	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	109644	56.11	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	112.22%	
35) Dibromofluoromethane		7.55	113	98477	47.73	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.46%	
50) Toluene-d8		10.06	98	401842	50.77	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.54%	
62) 4-Bromofluorobenzene		12.38	95	143721	52.97	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.94%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	16389	23.645	ug/l	89
20) Methylene Chloride	4.50	84	52821	29.915	ug/l	98
43) Isopropyl Acetate	8.13	43	14270	3.549	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121020\
Data File : VN065064.D
Acq On : 11 Dec 2020 3:00
Operator : JC/MD
Sample : PB133501ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133501ZHE#12

Quant Time: Dec 11 04:44:43 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 10 16:04:22 2020
Response via : Initial Calibration

