

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121020\
 Data File : VN065066.D
 Acq On : 11 Dec 2020 3:48
 Operator : JC/MD
 Sample : PB133501ZHE#14
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133501ZHE#14

Quant Time: Dec 11 04:59:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 10 16:04:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	160577	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	304834	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	330496	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	129569	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	108114	56.55	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	113.10%	
35) Dibromofluoromethane	7.55	113	98558	48.87	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	97.74%	
50) Toluene-d8	10.06	98	394907	51.04	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	102.08%	
62) 4-Bromofluorobenzene	12.37	95	144286	54.41	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	108.82%	

Target Compounds					Qvalue
16) Acetone	3.78	43	17117	25.243 ug/l	91
20) Methylene Chloride	4.50	84	56762	32.859 ug/l	99
43) Isopropyl Acetate	8.13	43	14381	3.660 ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121020\
Data File : VN065066.D
Acq On : 11 Dec 2020 3:48
Operator : JC/MD
Sample : PB133501ZHE#14
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133501ZHE#14

Quant Time: Dec 11 04:59:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 10 16:04:22 2020
Response via : Initial Calibration

