

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121020\
 Data File : VN065067.D
 Acq On : 11 Dec 2020 4:13
 Operator : JC/MD
 Sample : PB133501ZHE#15
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB133501ZHE#15

Quant Time: Dec 11 05:04:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 10 16:04:22 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	152348	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	294398	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	315212	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	125449	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	102885	56.72	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	113.44%	
35) Dibromofluoromethane		7.55	113	95017	48.79	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.58%	
50) Toluene-d8		10.06	98	374638	50.14	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.28%	
62) 4-Bromofluorobenzene		12.38	95	137535	53.70	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.40%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	16537	25.705	ug/l	94
20) Methylene Chloride	4.51	84	57150	34.870	ug/l	96
43) Isopropyl Acetate	8.13	43	16103	4.243	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121020\
Data File : VN065067.D
Acq On : 11 Dec 2020 4:13
Operator : JC/MD
Sample : PB133501ZHE#15
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB133501ZHE#15

Quant Time: Dec 11 05:04:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121020W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 10 16:04:22 2020
Response via : Initial Calibration

