

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
 Data File : VN069958.D
 Acq On : 10 Dec 2021 13:17
 Operator : JC/MD
 Sample : VN1210WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1210WBL01

Quant Time: Dec 11 05:18:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1106863	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2092817	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2407221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	980603	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	894588	52.346	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.700%
35) Dibromofluoromethane	8.021	113	660426	50.380	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.760%
50) Toluene-d8	10.440	98	2757006	52.378	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.760%
62) 4-Bromofluorobenzene	12.731	95	1173255	61.526	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	123.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
Data File : VN069958.D
Acq On : 10 Dec 2021 13:17
Operator : JC/MD
Sample : VN1210WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1210WBL01

Quant Time: Dec 11 05:18:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

