

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
 Data File : VN069963.D
 Acq On : 10 Dec 2021 17:15
 Operator : JC/MD
 Sample : M4970-02DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-19DL

Quant Time: Dec 11 05:21:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1092719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2087647	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2382694	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	962581	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	891034	52.813	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 105.620%			
35) Dibromofluoromethane	8.024	113	653453	49.971	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 99.940%			
50) Toluene-d8	10.443	98	2745864	52.295	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 104.600%			
62) 4-Bromofluorobenzene	12.733	95	1154082	60.671	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 121.340%			
Target Compounds						
11) Tert butyl alcohol	5.385	59	156725	49.143	ug/l	97
19) Methyl tert-butyl Ether	5.621	73	1601111	39.562	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
Data File : VN069963.D
Acq On : 10 Dec 2021 17:15
Operator : JC/MD
Sample : M4970-02DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-19DL

Quant Time: Dec 11 05:21:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

