

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
 Data File : VN069972.D
 Acq On : 10 Dec 2021 21:05
 Operator : JC/MD
 Sample : M4970-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-18

Quant Time: Dec 11 05:25:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1105193	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2103575	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2456390	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1052707	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	879670	51.551	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 103.100%	
35) Dibromofluoromethane	8.024	113	643162	48.812	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.620%	
50) Toluene-d8	10.443	98	2795873	52.845	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.680%	
62) 4-Bromofluorobenzene	12.731	95	1230370	64.191	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 128.380%	
Target Compounds						Qvalue
11) Tert butyl alcohol	5.382	59	867267	268.874	ug/l	98
16) Acetone	4.291	43	72969	7.109	ug/l	92
19) Methyl tert-butyl Ether	5.621	73	1865436	45.573	ug/l #	86
22) Diisopropyl ether	6.503	45	289337	6.240	ug/l #	90
39) Methylcyclohexane	9.461	83	15661	0.657	ug/l	92
40) Benzene	8.461	78	4792449	78.295	ug/l	98
52) Toluene	10.505	92	8043	0.215	ug/l	99
67) Ethyl Benzene	11.846	91	713909	7.887	ug/l	97
68) m/p-Xylenes	11.958	106	417602	12.670	ug/l	92
69) o-Xylene	12.283	106	12312	0.371	ug/l	79
73) Isopropylbenzene	12.581	105	103985	1.110	ug/l	99
78) n-propylbenzene	12.921	91	116013	1.113	ug/l	95
80) 1,3,5-Trimethylbenzene	13.058	105	82981	1.101	ug/l	96
84) 1,2,4-Trimethylbenzene	13.369	105	254020	3.538	ug/l	97
95) Naphthalene	15.509	128	79819m	5.413	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
Data File : VN069972.D
Acq On : 10 Dec 2021 21:05
Operator : JC/MD
Sample : M4970-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-18

Quant Time: Dec 11 05:25:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

