

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
 Data File : VN069978.D
 Acq On : 11 Dec 2021 00:28
 Operator : JC/MD
 Sample : VN1210WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1210WBL02

Quant Time: Dec 11 05:28:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1009446	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1911009	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2172643	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	870020	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	813475	52.193	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.380%
35) Dibromofluoromethane	8.024	113	598053	49.962	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.920%
50) Toluene-d8	10.443	98	2518040	52.389	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.780%
62) 4-Bromofluorobenzene	12.731	95	1048571	60.219	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	120.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
Data File : VN069978.D
Acq On : 11 Dec 2021 00:28
Operator : JC/MD
Sample : VN1210WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1210WBL02

Quant Time: Dec 11 05:28:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

