

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
 Data File : VN069983.D
 Acq On : 11 Dec 2021 3:26
 Operator : JC/MD
 Sample : PB141287ZHE#02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141287ZHE#02

Quant Time: Dec 11 05:30:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1073865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2091583	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2430993	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1015796	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	884790	53.363	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.720%	
35) Dibromofluoromethane	8.024	113	642631	49.051	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.100%	
50) Toluene-d8	10.443	98	2797744	53.183	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.360%	
62) 4-Bromofluorobenzene	12.733	95	1192534	62.574	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 125.140%	
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	67572	6.775	ug/l	92
20) Methylene Chloride	5.103	84	37087	2.797	ug/l #	84
25) 2-Butanone	7.348	43	17325	1.278	ug/l	90
43) Isopropyl Acetate	8.566	43	156665	3.572	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
Data File : VN069983.D
Acq On : 11 Dec 2021 3:26
Operator : JC/MD
Sample : PB141287ZHE#02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141287ZHE#02

Quant Time: Dec 11 05:30:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

