

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
 Data File : VN069984.D
 Acq On : 11 Dec 2021 3:52
 Operator : JC/MD
 Sample : PB141287ZHE#03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141287ZHE#03

Quant Time: Dec 11 05:31:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1127296	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2222742	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2592552	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	1081061	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	930933	53.485	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.980%
35) Dibromofluoromethane	8.024	113	678776	48.753	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.500%
50) Toluene-d8	10.443	98	2948197	52.736	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.480%
62) 4-Bromofluorobenzene	12.734	95	1273182	62.864	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	125.720%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	67194	6.418	ug/l	96
20) Methylene Chloride	5.095	84	38422	2.760	ug/l	92
43) Isopropyl Acetate	8.566	43	160587	3.446	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
Data File : VN069984.D
Acq On : 11 Dec 2021 3:52
Operator : JC/MD
Sample : PB141287ZHE#03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141287ZHE#03

Quant Time: Dec 11 05:31:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

