

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
 Data File : VN069989.D
 Acq On : 11 Dec 2021 6:00
 Operator : JC/MD
 Sample : M4990-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R3-7

Quant Time: Dec 11 08:31:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1096687	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2175363	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2519643	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1066283	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	913569	53.952	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	107.900%
35) Dibromofluoromethane	8.024	113	668842	49.086	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.180%
50) Toluene-d8	10.443	98	2890974	52.839	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.680%
62) 4-Bromofluorobenzene	12.734	95	1237621	62.439	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	124.880%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	71728	7.042	ug/l	95
18) Methyl Acetate	4.865	43	44457	1.446	ug/l	92
20) Methylene Chloride	5.098	84	53229	3.931	ug/l #	82
43) Isopropyl Acetate	8.566	43	172137	3.774	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
Data File : VN069989.D
Acq On : 11 Dec 2021 6:00
Operator : JC/MD
Sample : M4990-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R3-7

Quant Time: Dec 11 08:31:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

