

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
 Data File : VN069991.D
 Acq On : 11 Dec 2021 6:51
 Operator : JC/MD
 Sample : M4989-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P23-2

Quant Time: Dec 11 08:32:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1054070	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2066490	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2437173	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	1007857	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	865514	53.181	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.360%	
35) Dibromofluoromethane	8.024	113	643539	49.717	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.440%	
50) Toluene-d8	10.443	98	2770259	53.300	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.600%	
62) 4-Bromofluorobenzene	12.733	95	1182820	62.818	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 125.640%	
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	68901	7.038	ug/l	94
18) Methyl Acetate	4.867	43	42931	1.453	ug/l #	71
20) Methylene Chloride	5.101	84	52208	4.011	ug/l #	86
43) Isopropyl Acetate	8.566	43	157942	3.645	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
Data File : VN069991.D
Acq On : 11 Dec 2021 6:51
Operator : JC/MD
Sample : M4989-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P23-2

Quant Time: Dec 11 08:32:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

