

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
 Data File : VN069992.D
 Acq On : 11 Dec 2021 7:16
 Operator : JC/MD
 Sample : M4965-32
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FLOORS-COMP

Quant Time: Dec 11 08:32:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1071477	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2086367	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2420139	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	1046596	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	878313	53.091	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	106.180%
35) Dibromofluoromethane	8.024	113	644085	49.285	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.580%
50) Toluene-d8	10.443	98	2778917	52.957	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.920%
62) 4-Bromofluorobenzene	12.731	95	1192485	62.728	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	125.460%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	96619	9.709	ug/l	100
20) Methylene Chloride	5.098	84	42410	3.206	ug/l #	79
43) Isopropyl Acetate	8.563	43	197539	4.516	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
Data File : VN069992.D
Acq On : 11 Dec 2021 7:16
Operator : JC/MD
Sample : M4965-32
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FLOORS-COMP

Quant Time: Dec 11 08:32:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

