

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
 Data File : VN069993.D
 Acq On : 11 Dec 2021 7:42
 Operator : JC/MD
 Sample : M4970-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-24

Quant Time: Dec 11 08:33:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1067838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2076541	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2398791	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	959689	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	877866	53.245	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.480%	
35) Dibromofluoromethane	8.024	113	652379	50.156	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 100.320%	
50) Toluene-d8	10.443	98	2723440	52.146	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.300%	
62) 4-Bromofluorobenzene	12.731	95	1147080	60.625	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 121.260%	
Target Compounds						
16) Acetone	4.296	43	40240	4.057	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
Data File : VN069993.D
Acq On : 11 Dec 2021 7:42
Operator : JC/MD
Sample : M4970-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-24

Quant Time: Dec 11 08:33:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

