

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
 Data File : VN069994.D
 Acq On : 11 Dec 2021 8:07
 Operator : JC/MD
 Sample : M4970-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-30

Quant Time: Dec 11 08:33:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1049522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2048926	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2348213	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	945752	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	860772	53.119	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 106.240%	
35) Dibromofluoromethane	8.024	113	634792	49.462	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.920%	
50) Toluene-d8	10.443	98	2697770	52.350	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.700%	
62) 4-Bromofluorobenzene	12.733	95	1117306	59.847	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 119.700%	
Target Compounds						Qvalue
16) Acetone	4.299	43	30443	3.123	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
Data File : VN069994.D
Acq On : 11 Dec 2021 8:07
Operator : JC/MD
Sample : M4970-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-30

Quant Time: Dec 11 08:33:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

