

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
 Data File : VN069995.D
 Acq On : 11 Dec 2021 8:33
 Operator : JC/MD
 Sample : M4970-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-31

Quant Time: Dec 16 18:06:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1070511	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	2103360	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	2430462	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	1021955	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	898624	54.37	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.74%	
35) Dibromofluoromethane	8.024	113	655407	49.75	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 99.50%	
50) Toluene-d8	10.443	98	2765492	52.28	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.56%	
62) 4-Bromofluorobenzene	12.731	95	1214250	63.36	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 126.72%	
Target Compounds						
16) Acetone	4.293	43	41522	4.18	ug/l	96
19) Methyl tert-butyl Ether	5.618	73	172911	4.36	ug/l #	90
22) Diisopropyl ether	6.501	45	207111	4.61	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121021\
Data File : VN069995.D
Acq On : 11 Dec 2021 8:33
Operator : JC/MD
Sample : M4970-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-31

Quant Time: Dec 16 18:06:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

