

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121024\
 Data File : VN085157.D
 Acq On : 10 Dec 2024 11:15
 Operator : JC\MD
 Sample : VN1210WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1210WBL01

Quant Time: Dec 11 04:11:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	144184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	251448	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	221103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	86343	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	108935	53.041	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	106.080%	
35) Dibromofluoromethane	8.165	113	87621	51.673	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	103.340%	
50) Toluene-d8	10.565	98	299374	49.264	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.520%	
62) 4-Bromofluorobenzene	12.847	95	99708	43.875	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	87.740%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121024\
Data File : VN085157.D
Acq On : 10 Dec 2024 11:15
Operator : JC\MD
Sample : VN1210WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1210WBL01

Quant Time: Dec 11 04:11:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

