

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121024\
 Data File : VN085160.D
 Acq On : 10 Dec 2024 12:40
 Operator : JC\MD
 Sample : PB165471TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB165471TB

Quant Time: Dec 11 04:13:36 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	136968	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	247632	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	229506	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	91372	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	102411	52.491	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 104.980%		
35) Dibromofluoromethane	8.165	113	83481	49.990	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.980%		
50) Toluene-d8	10.565	98	306041	51.137	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.280%		
62) 4-Bromofluorobenzene	12.847	95	106258	47.477	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 94.960%		
Target Compounds						
16) Acetone	4.430	43	39468	69.460	ug/l	92
43) Isopropyl Acetate	8.688	43	118530	27.750	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121024\
Data File : VN085160.D
Acq On : 10 Dec 2024 12:40
Operator : JC\MD
Sample : PB165471TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB165471TB

Quant Time: Dec 11 04:13:36 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

