

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN121024\  
 Data File : VN085161.D  
 Acq On : 10 Dec 2024 13:04  
 Operator : JC\MD  
 Sample : P5174-02  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 ROLL-OFF-COMP

Quant Time: Dec 11 04:14:06 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N112224W.M  
 Quant Title : SW846 8260  
 QLast Update : Sat Nov 23 02:54:10 2024  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc   | Units  | Dev(Min) |
|-----------------------------|----------------|------|---------------------|--------|--------|----------|
| -----                       |                |      |                     |        |        |          |
| Internal Standards          |                |      |                     |        |        |          |
| 1) Pentafluorobenzene       | 8.224          | 168  | 155910              | 50.000 | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 9.100          | 114  | 285483              | 50.000 | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 11.865         | 117  | 262088              | 50.000 | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 13.788         | 152  | 101244              | 50.000 | ug/l   | 0.00     |
|                             |                |      |                     |        |        |          |
| System Monitoring Compounds |                |      |                     |        |        |          |
| 33) 1,2-Dichloroethane-d4   | 8.577          | 65   | 115082              | 51.819 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 74 - 125 |      | Recovery = 103.640% |        |        |          |
| 35) Dibromofluoromethane    | 8.165          | 113  | 94769               | 49.226 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 75 - 124 |      | Recovery = 98.460%  |        |        |          |
| 50) Toluene-d8              | 10.565         | 98   | 346674              | 50.246 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 86 - 113 |      | Recovery = 100.500% |        |        |          |
| 62) 4-Bromofluorobenzene    | 12.847         | 95   | 117683              | 45.610 | ug/l   | 0.00     |
| Spiked Amount 50.000        | Range 77 - 121 |      | Recovery = 91.220%  |        |        |          |
|                             |                |      |                     |        |        |          |
| Target Compounds            |                |      |                     |        |        |          |
| 16) Acetone                 | 4.436          | 43   | 15216               | 23.525 | ug/l   | 97       |
| 20) Methylene Chloride      | 5.277          | 84   | 2257                | 1.144  | ug/l # | 89       |
| 43) Isopropyl Acetate       | 8.688          | 43   | 119214              | 24.210 | ug/l # | 87       |
| -----                       |                |      |                     |        |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN121024\  
Data File : VN085161.D  
Acq On : 10 Dec 2024 13:04  
Operator : JC\MD  
Sample : P5174-02  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 8 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
ROLL-OFF-COMP

Quant Time: Dec 11 04:14:06 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N112224W.M  
Quant Title : SW846 8260  
QLast Update : Sat Nov 23 02:54:10 2024  
Response via : Initial Calibration

