

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121024\
 Data File : VN085167.D
 Acq On : 10 Dec 2024 15:31
 Operator : JC\MD
 Sample : P5219-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MR-CAM-013

Quant Time: Dec 11 04:16:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	135983	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	256711	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	235841	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	97141	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	104610	54.007	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.020%			
35) Dibromofluoromethane	8.165	113	86113	49.743	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.480%			
50) Toluene-d8	10.565	98	309938	49.957	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.920%			
62) 4-Bromofluorobenzene	12.847	95	108378	46.712	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 93.420%			
Target Compounds						
16) Acetone	4.430	43	56884	100.836	ug/l	94
43) Isopropyl Acetate	8.688	43	159881	36.107	ug/l #	81
52) Toluene	10.630	92	31643	7.207	ug/l	96
67) Ethyl Benzene	11.965	91	13660	1.574	ug/l	95
68) m/p-Xylenes	12.071	106	19178	5.805	ug/l	94
69) o-Xylene	12.394	106	7018	2.233	ug/l	99
70) Styrene	12.406	104	5265	1.007	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121024\
Data File : VN085167.D
Acq On : 10 Dec 2024 15:31
Operator : JC\MD
Sample : P5219-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MR-CAM-013

Quant Time: Dec 11 04:16:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

