

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
 Data File : VN088494.D
 Acq On : 10 Dec 2025 13:04
 Operator : JC\MD
 Sample : Q3813-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TB01-120825

Quant Time: Dec 11 02:06:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	225847	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	531810	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	534272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	252826	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	259278	60.942	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	121.880%
35) Dibromofluoromethane	8.147	113	180608	49.006	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.020%
50) Toluene-d8	10.541	98	618777	45.125	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.240%
62) 4-Bromofluorobenzene	12.829	95	228926	48.655	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.320%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
Data File : VN088494.D
Acq On : 10 Dec 2025 13:04
Operator : JC\MD
Sample : Q3813-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TB01-120825

Quant Time: Dec 11 02:06:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

