

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
 Data File : VN088503.D
 Acq On : 10 Dec 2025 16:10
 Operator : JC\MD
 Sample : Q3814-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-12

Quant Time: Dec 11 02:10:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	201767	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	481546	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	516525	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	224885	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	236141	62.128	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 124.260%	
35) Dibromofluoromethane	8.147	113	167365	50.153	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.300%	
50) Toluene-d8	10.541	98	609488	49.087	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.180%	
62) 4-Bromofluorobenzene	12.829	95	219170	51.444	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 102.880%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	166639	121.227	ug/l	99
20) Methylene Chloride	5.271	84	269961	91.793	ug/l	97
25) 2-Butanone	7.477	43	13983	6.328	ug/l #	84
43) Isopropyl Acetate	8.671	43	28232	2.729	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
Data File : VN088503.D
Acq On : 10 Dec 2025 16:10
Operator : JC\MD
Sample : Q3814-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-12

Quant Time: Dec 11 02:10:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

