

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
 Data File : VN088507.D
 Acq On : 10 Dec 2025 17:34
 Operator : JC\MD
 Sample : Q3826-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-13

Quant Time: Dec 11 02:12:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	221792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	518350	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	548520	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.771	152	250354	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	247931	59.340	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	118.680%	
35) Dibromofluoromethane	8.147	113	178460	49.681	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	99.360%	
50) Toluene-d8	10.541	98	631929	47.280	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	94.560%	
62) 4-Bromofluorobenzene	12.829	95	244881	53.398	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	106.800%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	207851	137.556	ug/l	99
20) Methylene Chloride	5.271	84	269545	83.377	ug/l	97
25) 2-Butanone	7.471	43	15086	6.210	ug/l	95
43) Isopropyl Acetate	8.671	43	23732	2.131	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
Data File : VN088507.D
Acq On : 10 Dec 2025 17:34
Operator : JC\MD
Sample : Q3826-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-13

Quant Time: Dec 11 02:12:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

