

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
 Data File : VN088508.D
 Acq On : 10 Dec 2025 17:54
 Operator : JC\MD
 Sample : Q3826-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-14

Quant Time: Dec 11 02:13:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	201740	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	476029	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	503228	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	226245	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	226107	59.496	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	119.000%	
35) Dibromofluoromethane	8.153	113	161998	49.107	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.220%	
50) Toluene-d8	10.541	98	571374	46.550	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	93.100%	
62) 4-Bromofluorobenzene	12.829	95	210265	49.926	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	99.860%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	180932	131.643	ug/l	95
20) Methylene Chloride	5.265	84	263791	89.707	ug/l	97
25) 2-Butanone	7.471	43	13603	6.156	ug/l #	88
43) Isopropyl Acetate	8.676	43	27471	2.686	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
Data File : VN088508.D
Acq On : 10 Dec 2025 17:54
Operator : JC\MD
Sample : Q3826-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-14

Quant Time: Dec 11 02:13:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

