

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
 Data File : VN088511.D
 Acq On : 10 Dec 2025 18:57
 Operator : JC\MD
 Sample : Q3826-10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Dec 11 02:14:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	193464	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	447561	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	478987	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	211639	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	211076	57.917	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 115.840%		
35) Dibromofluoromethane	8.153	113	151335	48.793	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 97.580%		
50) Toluene-d8	10.541	98	544475	47.180	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 94.360%		
62) 4-Bromofluorobenzene	12.829	95	198396	50.104	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 100.200%		
Target Compounds						
16) Acetone	4.424	43	195286	148.165	ug/l	99
20) Methylene Chloride	5.271	84	260437	92.355	ug/l	97
25) 2-Butanone	7.477	43	15116	7.134	ug/l #	79
43) Isopropyl Acetate	8.677	43	26102	2.714	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
Data File : VN088511.D
Acq On : 10 Dec 2025 18:57
Operator : JC\MD
Sample : Q3826-10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Dec 11 02:14:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

