

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
 Data File : VN088512.D
 Acq On : 10 Dec 2025 19:18
 Operator : JC\MD
 Sample : Q3826-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Dec 11 02:15:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	195745	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	471400	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	502812	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	230085	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	233211	63.245	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 126.480%#		
35) Dibromofluoromethane	8.147	113	164837	50.459	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 100.920%		
50) Toluene-d8	10.547	98	588788	48.440	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 96.880%		
62) 4-Bromofluorobenzene	12.829	95	218423	52.372	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 104.740%		
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	193193	144.869	ug/l	98
20) Methylene Chloride	5.265	84	267376	93.711	ug/l	97
25) 2-Butanone	7.477	43	13897	6.482	ug/l	99
43) Isopropyl Acetate	8.677	43	22778	2.249	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
Data File : VN088512.D
Acq On : 10 Dec 2025 19:18
Operator : JC\MD
Sample : Q3826-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Dec 11 02:15:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

