

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
 Data File : VN088513.D
 Acq On : 10 Dec 2025 19:38
 Operator : JC\MD
 Sample : Q3826-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Dec 11 02:15:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.212	168	209105	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	483731	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	523529	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	228541	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	230593	58.539	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	117.080%
35) Dibromofluoromethane	8.147	113	165913	49.493	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.980%
50) Toluene-d8	10.547	98	583238	46.760	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.520%
62) 4-Bromofluorobenzene	12.829	95	216603	50.612	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.220%
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	178501	125.300	ug/l	99
20) Methylene Chloride	5.265	84	262355	86.076	ug/l	96
25) 2-Butanone	7.477	43	17154	7.490	ug/l #	81
43) Isopropyl Acetate	8.671	43	28409	2.733	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121025\
Data File : VN088513.D
Acq On : 10 Dec 2025 19:38
Operator : JC\MD
Sample : Q3826-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Dec 11 02:15:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

