

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121118\
 Data File : VN052796.D
 Acq On : 11 Dec 2018 12:35
 Operator : MD\SY
 Sample : PB115504ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115504ZHE#01

Quant Time: Dec 12 01:17:13 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1244205	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1939313	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1759971	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	724571	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	644234	47.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.12%	
35) Dibromofluoromethane	7.59	113	543091	45.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.84%	
50) Toluene-d8	10.09	98	2382355	51.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.38%	
62) 4-Bromofluorobenzene	12.40	95	790781	50.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.14%	
Target Compounds						
16) Acetone	3.82	43	10536	2.949	ug/l	99
20) Methylene Chloride	4.55	84	17544	1.255	ug/l	99
43) Isopropyl Acetate	8.17	43	61233	3.237	ug/l #	77
95) Naphthalene	15.13	128	3941	3.563	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121118\
Data File : VN052796.D
Acq On : 11 Dec 2018 12:35
Operator : MD\SY
Sample : PB115504ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115504ZHE#01

Quant Time: Dec 12 01:17:13 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

