

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121118\
 Data File : VN052798.D
 Acq On : 11 Dec 2018 13:28
 Operator : MD\SY
 Sample : PB115504ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115504ZHE#03

Quant Time: Dec 12 01:22:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1168378	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1827587	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1647233	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	699280	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	616854	48.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.98%	
35) Dibromofluoromethane	7.59	113	509780	45.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.50%	
50) Toluene-d8	10.09	98	2243123	51.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.30%	
62) 4-Bromofluorobenzene	12.40	95	749891	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
Target Compounds						
16) Acetone	3.82	43	11294	3.366	ug/l	97
20) Methylene Chloride	4.56	84	19377	1.476	ug/l	94
43) Isopropyl Acetate	8.17	43	67341	3.778	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121118\
Data File : VN052798.D
Acq On : 11 Dec 2018 13:28
Operator : MD\SY
Sample : PB115504ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115504ZHE#03

Quant Time: Dec 12 01:22:38 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

