

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121118\
 Data File : VN052805.D
 Acq On : 11 Dec 2018 16:35
 Operator : MD\SY
 Sample : PB115504ZHE#10
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115504ZHE#10

Quant Time: Dec 12 02:17:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1219351	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1919479	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1737373	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	740021	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	653496	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
35) Dibromofluoromethane	7.59	113	534867	45.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.40%	
50) Toluene-d8	10.09	98	2353006	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
62) 4-Bromofluorobenzene	12.40	95	786241	50.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.60%	
Target Compounds						
16) Acetone	3.82	43	17428	4.977	ug/l	100
43) Isopropyl Acetate	8.17	43	86405	4.615	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121118\
Data File : VN052805.D
Acq On : 11 Dec 2018 16:35
Operator : MD\SY
Sample : PB115504ZHE#10
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115504ZHE#10

Quant Time: Dec 12 02:17:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

