

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121118\
 Data File : VN052807.D
 Acq On : 11 Dec 2018 17:29
 Operator : MD\SY
 Sample : PB115504ZHE#12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 18 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115504ZHE#12

Quant Time: Dec 12 02:21:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1186822	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1850249	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1672846	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	691433	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	626016	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
35) Dibromofluoromethane	7.59	113	526132	46.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.24%	
50) Toluene-d8	10.09	98	2278069	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
62) 4-Bromofluorobenzene	12.40	95	744343	49.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.78%	
Target Compounds						
16) Acetone	3.82	43	14900	4.372	ug/l #	88
43) Isopropyl Acetate	8.17	43	90782	5.031	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121118\
Data File : VN052807.D
Acq On : 11 Dec 2018 17:29
Operator : MD\SY
Sample : PB115504ZHE#12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 18 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115504ZHE#12

Quant Time: Dec 12 02:21:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

