

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121119\
 Data File : VN059563.D
 Acq On : 11 Dec 2019 15:46
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 12 02:32:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 10 01:09:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	247083	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	395346	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	355953	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142475	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	175778	50.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.80%	
35) Dibromofluoromethane	7.57	113	124204	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
50) Toluene-d8	10.08	98	490170	50.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.78%	
62) 4-Bromofluorobenzene	12.40	95	162198	45.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.98%	
Target Compounds						
16) Acetone	3.79	43	52892	32.788	ug/l	96
25) 2-Butanone	6.82	43	18579	8.556	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121119\
Data File : VN059563.D
Acq On : 11 Dec 2019 15:46
Operator : JC/SP
Sample : IBLK
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 12 02:32:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120919W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 10 01:09:52 2019
Response via : Initial Calibration

