

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121121\
 Data File : VN070001.D
 Acq On : 11 Dec 2021 11:51
 Operator : JC/MD
 Sample : VN1211MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1211MBL01

Quant Time: Dec 12 16:17:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1000380	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1901998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2148427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	806451	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	807501	52.279	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.560%
35) Dibromofluoromethane	8.024	113	595786	50.008	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.020%
50) Toluene-d8	10.443	98	2514846	52.571	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.140%
62) 4-Bromofluorobenzene	12.734	95	1011469	58.364	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	116.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121121\
Data File : VN070001.D
Acq On : 11 Dec 2021 11:51
Operator : JC/MD
Sample : VN1211MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1211MBL01

Quant Time: Dec 12 16:17:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

