

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121121\
 Data File : VN070010.D
 Acq On : 11 Dec 2021 16:02
 Operator : JC/MD
 Sample : M4873-13
 Misc : 5.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 A6-20-7.5

Quant Time: Dec 12 16:22:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1235343	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	2328620	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2542273	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	1233614	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	974881	51.11	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 102.22%	
35) Dibromofluoromethane	8.016	113	699762	47.97	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 95.94%	
50) Toluene-d8	10.440	98	3052529	52.12	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.24%	
62) 4-Bromofluorobenzene	12.734	95	1312355	61.85	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 123.70%	
Target Compounds						
95) Naphthalene	15.509	128	34536	4.51	ug/l	Qvalue # 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121121\
Data File : VN070010.D
Acq On : 11 Dec 2021 16:02
Operator : JC/MD
Sample : M4873-13
Misc : 5.99g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
A6-20-7.5

Quant Time: Dec 12 16:22:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

