

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121121\
 Data File : VN070011.D
 Acq On : 11 Dec 2021 16:27
 Operator : JC/MD
 Sample : M4873-06
 Misc : 6.93g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 A2-7-6.5

Quant Time: Dec 12 16:22:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1154451	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	2192102	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2299207	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	1117871	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	909055	51.00	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	102.00%
35) Dibromofluoromethane	8.018	113	652374	47.51	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	95.02%
50) Toluene-d8	10.440	98	2798490	50.76	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.52%
62) 4-Bromofluorobenzene	12.733	95	1181551	59.15	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	118.30%
Target Compounds						
					Qvalue	
19) Methyl tert-butyl Ether	5.615	73	2582700	60.40	ug/l	95
39) Methylcyclohexane	9.453	83	25501	1.03	ug/l #	87
40) Benzene	8.453	78	365712	5.73	ug/l	98
67) Ethyl Benzene	11.846	91	189375	2.24	ug/l	98
68) m/p-Xylenes	11.950	106	73475	2.38	ug/l	95
78) n-propylbenzene	12.924	91	104622	0.95	ug/l	95
80) 1,3,5-Trimethylbenzene	13.058	105	62101	0.78	ug/l	96
84) 1,2,4-Trimethylbenzene	13.369	105	202424	2.66	ug/l	98
95) Naphthalene	15.507	128	124194	6.08	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121121\
Data File : VN070011.D
Acq On : 11 Dec 2021 16:27
Operator : JC/MD
Sample : M4873-06
Misc : 6.93g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
A2-7-6.5

Quant Time: Dec 12 16:22:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

