

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121121\
 Data File : VN070021.D
 Acq On : 11 Dec 2021 20:44
 Operator : JC/MD
 Sample : M4954-05ME
 Misc : 6.68g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-5(7.5-8)ME

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 12/15/2021
 Supervised By :Semsettin Yesilyurt 12/15/2021

Quant Time: Dec 12 16:27:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1228925	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	2287020	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2533731	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	1396509	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	950290	50.082	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.160%			
35) Dibromofluoromethane	8.019	113	679564	47.438	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 94.880%			
50) Toluene-d8	10.440	98	3079814	53.542	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.080%			
62) 4-Bromofluorobenzene	12.734	95	1430851	68.663	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 137.320%#			
Target Compounds					Qvalue	
31) Cyclohexane	8.083	56	122560m	4.641	ug/l	
39) Methylcyclohexane	9.456	83	518641	20.027	ug/l	89
67) Ethyl Benzene	11.846	91	674609	7.225	ug/l	95
68) m/p-Xylenes	11.953	106	896034	26.356	ug/l	93
69) o-Xylene	12.283	106	37745	1.102	ug/l	68
73) Isopropylbenzene	12.581	105	604062	4.862	ug/l	98
78) n-propylbenzene	12.921	91	1362580	9.855	ug/l	95
80) 1,3,5-Trimethylbenzene	13.061	105	2801316	28.011	ug/l	97
84) 1,2,4-Trimethylbenzene	13.369	105	5120434	53.762	ug/l	97
85) sec-Butylbenzene	13.503	105	1405989	11.705	ug/l	88
86) p-Isopropyltoluene	13.616	119	1021273	10.878	ug/l	96
89) n-Butylbenzene	13.943	91	1151470	15.119	ug/l #	72
95) Naphthalene	15.504	128	1277696	21.215	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN121121\
Data File : VN070021.D
Acq On : 11 Dec 2021 20:44
Operator : JC/MD
Sample : M4954-05ME
Misc : 6.68g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-5(7.5-8)ME

Quant Time: Dec 12 16:27:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 12/15/2021
Supervised By :Semsettin Yesilyurt 12/15/2021

