

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121218\
 Data File : VN052841.D
 Acq On : 12 Dec 2018 21:33
 Operator : MD\SY
 Sample : PB115564ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115564ZHE#05

Quant Time: Dec 13 04:48:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1154598	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1817397	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1634904	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	681963	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	613577	49.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.96%	
35) Dibromofluoromethane	7.59	113	522628	51.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.04%	
50) Toluene-d8	10.09	98	2226831	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
62) 4-Bromofluorobenzene	12.40	95	735097	49.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.64%	
Target Compounds						
16) Acetone	3.82	43	15413	5.590	ug/l	94
20) Methylene Chloride	4.56	84	21245	1.712	ug/l	97
43) Isopropyl Acetate	8.17	43	94535	4.586	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121218\
Data File : VN052841.D
Acq On : 12 Dec 2018 21:33
Operator : MD\SY
Sample : PB115564ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115564ZHE#05

Quant Time: Dec 13 04:48:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

