

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121218\
 Data File : VN052847.D
 Acq On : 13 Dec 2018 00:14
 Operator : MD\SY
 Sample : PB115564ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115564ZHE#11

Quant Time: Dec 13 05:01:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1131058	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1777510	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1631344	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	688232	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	603385	50.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.34%	
35) Dibromofluoromethane	7.59	113	510729	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	
50) Toluene-d8	10.09	98	2191628	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
62) 4-Bromofluorobenzene	12.40	95	733246	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
Target Compounds						
16) Acetone	3.82	43	18014	6.669	ug/l	100
20) Methylene Chloride	4.55	84	26338	2.167	ug/l	94
43) Isopropyl Acetate	8.17	43	79687	3.953	ug/l #	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121218\
Data File : VN052847.D
Acq On : 13 Dec 2018 00:14
Operator : MD\SY
Sample : PB115564ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115564ZHE#11

Quant Time: Dec 13 05:01:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration

