

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121218\
 Data File : VN052837.D
 Acq On : 12 Dec 2018 19:47
 Operator : MD\SY
 Sample : PB115564ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115564ZHE#01

Quant Time: Dec 13 04:37:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 13 01:08:35 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1149500	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1796287	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1655347	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	698202	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	607332	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
35) Dibromofluoromethane	7.59	113	516484	51.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.02%	
50) Toluene-d8	10.09	98	2221215	51.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.18%	
62) 4-Bromofluorobenzene	12.40	95	745911	50.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.28%	
Target Compounds						
16) Acetone	3.82	43	17162	6.252	ug/l	97
20) Methylene Chloride	4.55	84	14863	1.203	ug/l	94
43) Isopropyl Acetate	8.17	43	73547	3.610	ug/l #	80


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121218\
Data File : VN052837.D
Acq On : 12 Dec 2018 19:47
Operator : MD\SY
Sample : PB115564ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115564ZHE#01

Quant Time: Dec 13 04:37:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N121218W.M
Quant Title : SW846 8260
QLast Update : Thu Dec 13 01:08:35 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1833

Delta R.T. 0.00 min

Lab File: VN052837.D

Acq: 12 Dec 2018 19:47

Instrument :

MSVOA_N

ClientSampleId :


PB115564ZHE#01

Tot Ion:168 Resp: 1149500

Ion Ratio Lower Upper

168 100

99 50.2 40.2 60.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VN

7.67

500000

400000



300000

200000

100000

0

Time-->



#16

Acetone

Concen: 6.252 ug/l

RT: 3.82 min Scan# 636

Delta R.T. 0.00 min

Lab File: VN052837.D


Acq: 12 Dec 2018 19:47

Tgt Ion: 43 Resp: 17162

Ion Ratio Lower Upper

43 100

58 29.8 25.4 38.0



Abundance Ion 43.00 (42.70 to 43.70): VN

Ion 58.00 (57.70 to 58.70): VN

3.82


6000


4000

2000

0

Time-->









#20
Methvlene Chloride
Concen: 1.203 ug/l
RT: 4.55 min Scan# 864
Delta R.T. -0.00 min
Lab File: VN052837.D
Acq: 12 Dec 2018 19:47




Instrument :
MSVOA_N
ClientSampleId :
PB115564ZHE#01


Tot Ion: 84 Resp: 14863
Ion Ratio Lower Upper
84 100
49 111.8 96.8 145.2
51 38.6 29.9 44.9
86 60.6 51.8 77.6



#33
1,2-Dichloroethane-d4
Concen: N.D. ug/l
RT: 8.03 min Scan# 1945
Delta R.T. 0.00 min
Lab File: VN052837.D
Acq: 12 Dec 2018 19:47

Tgt Ion: 65 Resp: 607332
Ion Ratio Lower Upper
65 100
67 54.4 0.0 110.4









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN052837.D
 Acq: 12 Dec 2018 19:47

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115564ZHE#01

Tot Ion:114 Resp: 1796287
 Ion Ratio Lower Upper
 114 100
 63 19.7 0.0 39.8
 88 15.5 0.0 31.0




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VN
 Ion 87.90 (87.60 to 88.60): VN






#35
 Dibromofluoromethane
 Concen: 50.000 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN052837.D
 Acq: 12 Dec 2018 19:47

Tgt Ion:113 Resp: 516484
 Ion Ratio Lower Upper
 113 100
 111 102.6 83.0 124.4
 192 21.9 17.5 26.3



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 3.610 ug/l

RT: 8.17 min Scan# 1989

Delta R.T. 0.01 min

Lab File: VN052837.D

Acq: 12 Dec 2018 19:47

Instrument :

MSVOA_N

ClientSampleId :

PB115564ZHE#01


Tot Ion: 43 Resp: 73547

Ion Ratio Lower Upper

43 100

61 14.8 22.9 34.3#




87 10.0 10.6 15.8#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 61.00 (60.70 to 61.70): VN0

Ion 87.00 (86.70 to 87.70): VN0



#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587

Delta R.T. 0.00 min

Lab File: VN052837.D


Acq: 12 Dec 2018 19:47

Tgt Ion: 98 Resp: 2221215

Ion Ratio Lower Upper



98 100


100 63.9 51.6 77.4



Abundance Ion 98.00 (97.70 to 98.70): VN0

Ion 100.00 (99.70 to 100.70): VN0









#62
4-Bromofluorobenzene
Concen: Below ug/l
RT: 12.40 min Scan# 3306
Delta R.T. 0.00 min
Lab File: VN052837.D
Acq: 12 Dec 2018 19:47




Instrument :
MSVOA_N
ClientSampleId :
PB115564ZHE#01


Tot Ion: 95 Resp: 745911
Ion Ratio Lower Upper
95 100
174 89.6 0.0 178.8
176 86.0 0.0 173.8



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.41 min Scan# 2996
Delta R.T. 0.00 min
Lab File: VN052837.D
Acq: 12 Dec 2018 19:47

Tgt Ion: 117 Resp: 1655347
Ion Ratio Lower Upper
117 100
82 53.4 42.8 64.2
119 31.6 25.5 38.3





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.35 min Scan# 3599

Delta R.T. 0.00 min

Lab File: VN052837.D

Acq: 12 Dec 2018 19:47

Instrument :

MSVOA_N

ClientSampleId :

PB115564ZHE#01


Tot Ion:152 Resp: 698202

Ion Ratio Lower Upper

152 100

115 56.4 28.3 85.0

150 157.5 0.0 347.8



Abundance

Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V

