

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN121219\
 Data File : VN059569.D
 Acq On : 12 Dec 2019 12:44
 Operator : JC/SP
 Sample : VN1212MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1212MBL01

Quant Time: Dec 13 04:36:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 10 01:09:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	234316	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	367057	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	321148	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	119217	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.00	65	163055	49.79	ug/l	0.00
Spiked Amount			Recovery	=	99.58%	
35) Dibromofluoromethane	7.56	113	115260	52.62	ug/l	0.00
Spiked Amount			Recovery	=	105.24%	
50) Toluene-d8	10.08	98	452686	50.62	ug/l	0.00
Spiked Amount			Recovery	=	101.24%	
62) 4-Bromofluorobenzene	12.40	95	143370	43.31	ug/l	0.00
Spiked Amount			Recovery	=	86.62%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN121219\
Data File : VN059569.D
Acq On : 12 Dec 2019 12:44
Operator : JC/SP
Sample : VN1212MBL01
Misc : 5.00µL/10mL/100µL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1212MBL01

Quant Time: Dec 13 04:36:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120919W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 10 01:09:52 2019
Response via : Initial Calibration

