

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN121219\
 Data File : VN059570.D
 Acq On : 12 Dec 2019 13:18
 Operator : JC/SP
 Sample : VN1212WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1212WBL01

Quant Time: Dec 13 04:37:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120919W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 10 01:09:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	248194	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	399331	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	357686	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142921	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	178335	51.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.82%	
35) Dibromofluoromethane	7.57	113	125293	52.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.16%	
50) Toluene-d8	10.08	98	493842	50.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.52%	
62) 4-Bromofluorobenzene	12.40	95	163854	45.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.00%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN121219\
Data File : VN059570.D
Acq On : 12 Dec 2019 13:18
Operator : JC/SP
Sample : VN1212WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1212WBL01

Quant Time: Dec 13 04:37:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120919W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 10 01:09:52 2019
Response via : Initial Calibration

